

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021220\
 Data File : VN060077.D
 Acq On : 12 Feb 2020 16:44
 Operator : JC/MD
 Sample : L1491-04MEDL 10X
 Misc : 11.59g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-2A-(13.5-15.5)MEDL

Quant Time: Feb 13 01:17:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	199962	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	317741	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	294946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	132523	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	115808	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
35) Dibromofluoromethane	7.56	113	95629	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
50) Toluene-d8	10.08	98	381952	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
62) 4-Bromofluorobenzene	12.40	95	136727	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
Target Compounds						
39) Methylcyclohexane	9.06	83	2640	0.723	ug/l #	88
78) n-propylbenzene	12.59	91	7177	0.637	ug/l	98
85) sec-Butylbenzene	13.17	105	7281	0.755	ug/l	94
89) n-Butylbenzene	13.61	91	6319	0.823	ug/l	90
95) Naphthalene	15.13	128	13493	1.998	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021220\
Data File : VN060077.D
Acq On : 12 Feb 2020 16:44
Operator : JC/MD
Sample : L1491-04MEDL 10X
Misc : 11.59g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-2A-(13.5-15.5)MEDL

Quant Time: Feb 13 01:17:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

