

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
 Data File : VN065834.D
 Acq On : 12 Feb 2021 13:26
 Operator : JC/MD
 Sample : VN0212WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0212WBL02

Quant Time: Feb 13 00:47:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	120441	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	241119	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	243455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	86311	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	75127	47.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.82%
35) Dibromofluoromethane	7.547	113	69744	48.81	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.62%
50) Toluene-d8	10.055	98	307715	54.79	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.58%
62) 4-Bromofluorobenzene	12.370	95	116468	57.59	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	115.18%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
Data File : VN065834.D
Acq On : 12 Feb 2021 13:26
Operator : JC/MD
Sample : VN0212WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0212WBL02

Quant Time: Feb 13 00:47:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

