

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
 Data File : VN065845.D
 Acq On : 12 Feb 2021 17:59
 Operator : JC/MD
 Sample : PB134572ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134572ZHE#02

Manual Integrations
 APPROVED

MMDadoda
 2/15/2021 1:32:16 PM

Quant Time: Feb 13 02:38:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	6.650	168	126272m	50.00	ug/l	-0.97
34) 1,4-Difluorobenzene	7.958	114	249819m	50.00	ug/l	-0.59
63) Chlorobenzene-d5	11.254	117	262356	50.00	ug/l	-0.12
72) 1,4-Dichlorobenzene-d4	13.286	152	101870	50.00	ug/l	-0.03
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.155	65	78588m	47.31	ug/l	-0.83
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.62%		
35) Dibromofluoromethane	6.502	113	72777m	49.16	ug/l	-1.05
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.32%		
50) Toluene-d8	9.788	98	326186m	56.05	ug/l	-0.27
Spiked Amount 50.000	Range 65 - 126		Recovery =	112.10%		
62) 4-Bromofluorobenzene	12.312	95	126825	60.52	ug/l	-0.06
Spiked Amount 50.000	Range 58 - 135		Recovery =	121.04%		
Target Compounds						Qvalue
16) Acetone	1.788	43	34921m	54.00	ug/l	
20) Methylene Chloride	2.505	84	12723m	7.10	ug/l	
43) Isopropyl Acetate	8.855	43	775947m	181.44	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
Data File : VN065845.D
Acq On : 12 Feb 2021 17:59
Operator : JC/MD
Sample : PB134572ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134572ZHE#02

Manual Integrations
APPROVED

MMDadoda
2/15/2021 1:32:16 PM

Quant Time: Feb 13 02:38:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

