

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
 Data File : VN065846.D
 Acq On : 12 Feb 2021 18:22
 Operator : JC/MD
 Sample : PB134572ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134572ZHE#03

Quant Time: Feb 13 00:50:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	125696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	248679	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	264826	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	104569	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	79480	48.06	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.12%
35) Dibromofluoromethane	7.550	113	72504	49.20	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.40%
50) Toluene-d8	10.058	98	331411	57.21	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	114.42%
62) 4-Bromofluorobenzene	12.373	95	128546	61.63	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	123.26%
Target Compounds						
					Qvalue	
16) Acetone	3.779	43	25524	39.65	ug/l	89
20) Methylene Chloride	4.499	84	12491	7.01	ug/l	98
43) Isopropyl Acetate	8.132	43	13371	3.14	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
Data File : VN065846.D
Acq On : 12 Feb 2021 18:22
Operator : JC/MD
Sample : PB134572ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134572ZHE#03

Quant Time: Feb 13 00:50:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

