

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
 Data File : VN065850.D
 Acq On : 12 Feb 2021 20:01
 Operator : JC/MD
 Sample : M1382-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-VPB-RW6-GW-438-440

Quant Time: Feb 13 00:51:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	118203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	233931	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	241588	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	91790	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	72940	46.90	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.80%
35) Dibromofluoromethane	7.547	113	68640	49.51	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.02%
50) Toluene-d8	10.058	98	298856	54.84	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.68%
62) 4-Bromofluorobenzene	12.373	95	114798	58.50	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
Data File : VN065850.D
Acq On : 12 Feb 2021 20:01
Operator : JC/MD
Sample : M1382-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-VPB-RW6-GW-438-440

Quant Time: Feb 13 00:51:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

