

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
 Data File : VN065860.D
 Acq On : 13 Feb 2021 00:32
 Operator : JC/MD
 Sample : M1398-04MEDL 10X
 Misc : 3.44g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VCP-SS-4-SOILMEDL

Quant Time: Feb 13 03:39:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	116078	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	226435	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	241747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	109068	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	68913	45.13	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 90.26%			
35) Dibromofluoromethane	7.547	113	65308	48.67	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.34%			
50) Toluene-d8	10.055	98	297472	56.40	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 112.80%			
62) 4-Bromofluorobenzene	12.373	95	123193	64.86	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 129.72%			
Target Compounds						Qvalue
67) Ethyl Benzene	11.479	91	16892	1.78	ug/l #	88
68) m/p-Xylenes	11.589	106	17792	5.17	ug/l	86
73) Isopropylbenzene	12.219	105	16270	1.70	ug/l #	68
78) n-propylbenzene	12.560	91	48412	4.30	ug/l	98
80) 1,3,5-Trimethylbenzene	12.701	105	53464	6.58	ug/l	95
84) 1,2,4-Trimethylbenzene	13.010	105	227885	28.10	ug/l	100
85) sec-Butylbenzene	13.142	105	38079	4.16	ug/l	89
86) p-Isopropyltoluene	13.257	119	21983	2.72	ug/l	98
89) n-Butylbenzene	13.582	91	33843	4.66	ug/l #	74
95) Naphthalene	15.097	128	41908	8.97	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021221\
Data File : VN065860.D
Acq On : 13 Feb 2021 00:32
Operator : JC/MD
Sample : M1398-04MEDL 10X
Misc : 3.44g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VCP-SS-4-SOILMEDL

Quant Time: Feb 13 03:39:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

