

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021224\
 Data File : VN080999.D
 Acq On : 12 Feb 2024 11:33
 Operator : JC\MD
 Sample : VN0212WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0212WBL02

Quant Time: Feb 13 03:17:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	556967	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	950998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	800947	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	326833	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	303995	43.266	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.540%
35) Dibromofluoromethane	8.171	113	285700	55.986	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.980%
50) Toluene-d8	10.571	98	924318	50.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.160%
62) 4-Bromofluorobenzene	12.847	95	300968	44.884	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021224\
Data File : VN080999.D
Acq On : 12 Feb 2024 11:33
Operator : JC\MD
Sample : VN0212WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0212WBL02

Quant Time: Feb 13 03:17:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

