

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021224\
 Data File : VN081003.D
 Acq On : 12 Feb 2024 14:09
 Operator : JC\MD
 Sample : PB158904TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158904TB

Quant Time: Feb 13 03:20:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	415497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	767502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	693283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	309953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	239728	45.736	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.480%
35) Dibromofluoromethane	8.171	113	212392	51.571	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.140%
50) Toluene-d8	10.571	98	723843	48.591	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.180%
62) 4-Bromofluorobenzene	12.853	95	278514	51.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.940%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	46251	26.821	ug/l #	81
20) Methylene Chloride	5.259	84	15790	3.022	ug/l #	61
37) Ethyl Acetate	7.565	43	8431	1.819	ug/l #	90
43) Isopropyl Acetate	8.688	43	272229	31.933	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021224\
Data File : VN081003.D
Acq On : 12 Feb 2024 14:09
Operator : JC\MD
Sample : PB158904TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158904TB

Quant Time: Feb 13 03:20:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

