

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021224\
 Data File : VN081005.D
 Acq On : 12 Feb 2024 14:57
 Operator : JC\MD
 Sample : PB158948TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158948TB

Quant Time: Feb 13 03:21:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	426749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	785444	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	722638	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	303276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	228447	42.435	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.860%
35) Dibromofluoromethane	8.171	113	205665	48.797	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.600%
50) Toluene-d8	10.571	98	702404	46.075	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.140%
62) 4-Bromofluorobenzene	12.853	95	265370	47.917	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.840%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	25494	14.394	ug/l	94
20) Methylene Chloride	5.277	84	12077	2.251	ug/l	# 78
37) Ethyl Acetate	7.571	43	7936	1.673	ug/l	# 94
43) Isopropyl Acetate	8.694	43	275392	31.566	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021224\
Data File : VN081005.D
Acq On : 12 Feb 2024 14:57
Operator : JC\MD
Sample : PB158948TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158948TB

Quant Time: Feb 13 03:21:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

