

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053747.D
 Acq On : 13 Feb 2019 11:30
 Operator : JC/SP
 Sample : VN0213WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0213WBL01

Quant Time: Feb 14 00:37:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	627348	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1072596	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	983882	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	384492	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	345624	54.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.84%	
35) Dibromofluoromethane	7.59	113	345669	51.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.00%	
50) Toluene-d8	10.09	98	1299116	51.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.10%	
62) 4-Bromofluorobenzene	12.40	95	418824	49.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.70%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053747.D
Acq On : 13 Feb 2019 11:30
Operator : JC/SP
Sample : VN0213WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0213WBL01

Quant Time: Feb 14 00:37:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

