

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053749.D
 Acq On : 13 Feb 2019 12:34
 Operator : JC/SP
 Sample : PB117063TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117063TB

Quant Time: Feb 14 00:45:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	677814	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1159114	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1075585	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	420406	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	373472	54.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.84%	
35) Dibromofluoromethane	7.59	113	375571	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	10.09	98	1416686	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
62) 4-Bromofluorobenzene	12.40	95	459914	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
Target Compounds						
16) Acetone	3.82	43	13246	7.594	ug/l	94
20) Methylene Chloride	4.55	84	13430	1.835	ug/l	94
25) 2-Butanone	6.85	43	4848	1.949	ug/l	96
43) Isopropyl Acetate	8.17	43	35555	3.155	ug/l #	81
95) Naphthalene	15.14	128	3768	1.933	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053749.D
Acq On : 13 Feb 2019 12:34
Operator : JC/SP
Sample : PB117063TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117063TB

Quant Time: Feb 14 00:45:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

