

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053753.D
 Acq On : 13 Feb 2019 14:21
 Operator : JC/SP
 Sample : K1343-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-021119-D

Quant Time: Feb 14 00:54:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	708597	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1166251	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1082472	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	434942	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	377684	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
35) Dibromofluoromethane	7.59	113	371754	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
50) Toluene-d8	10.09	98	1418945	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	12.40	95	462550	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
Target Compounds						
16) Acetone	3.82	43	15863	8.700	ug/l	96
20) Methylene Chloride	4.56	84	75959	9.930	ug/l	99
43) Isopropyl Acetate	8.17	43	29415	2.594	ug/l #	83
95) Naphthalene	15.13	128	90158	6.291	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053753.D
Acq On : 13 Feb 2019 14:21
Operator : JC/SP
Sample : K1343-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-021119-D

Quant Time: Feb 14 00:54:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

