

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053754.D
 Acq On : 13 Feb 2019 14:48
 Operator : JC/SP
 Sample : K1343-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-021119-E

Quant Time: Feb 14 00:56:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	693634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1165694	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1080668	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	442340	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	376691	54.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.26%	
35) Dibromofluoromethane	7.59	113	369392	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
50) Toluene-d8	10.09	98	1413756	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
62) 4-Bromofluorobenzene	12.40	95	467867	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
Target Compounds						
16) Acetone	3.82	43	17608	9.865	ug/l	98
20) Methylene Chloride	4.55	84	101827	13.598	ug/l	98
43) Isopropyl Acetate	8.17	43	34249	3.022	ug/l #	78
95) Naphthalene	15.13	128	50314	4.236	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053754.D
Acq On : 13 Feb 2019 14:48
Operator : JC/SP
Sample : K1343-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-021119-E

Quant Time: Feb 14 00:56:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

