

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053756.D
 Acq On : 13 Feb 2019 15:42
 Operator : JC/SP
 Sample : K1343-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-021119-G

Quant Time: Feb 14 01:03:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	714676	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1190331	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1103439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	453946	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	384617	53.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.30%	
35) Dibromofluoromethane	7.59	113	380056	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
50) Toluene-d8	10.09	98	1439114	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	12.40	95	501916	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
Target Compounds						
16) Acetone	3.82	43	15465	8.409	ug/l	100
20) Methylene Chloride	4.56	84	444680	57.635	ug/l	99
43) Isopropyl Acetate	8.17	43	32626	2.819	ug/l #	80
95) Naphthalene	15.13	128	17960	2.606	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053756.D
Acq On : 13 Feb 2019 15:42
Operator : JC/SP
Sample : K1343-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-021119-G

Quant Time: Feb 14 01:03:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

