

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053758.D
 Acq On : 13 Feb 2019 16:35
 Operator : JC/SP
 Sample : K1343-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-021119-I

Quant Time: Feb 14 01:13:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	705093	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1171583	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1097073	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	448675	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	378676	53.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.06%	
35) Dibromofluoromethane	7.59	113	366823	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
50) Toluene-d8	10.09	98	1433117	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
62) 4-Bromofluorobenzene	12.40	95	472545	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
Target Compounds						
16) Acetone	3.82	43	16619	9.160	ug/l	93
20) Methylene Chloride	4.55	84	96296	12.651	ug/l	92
40) Benzene	8.04	78	42984	1.320	ug/l	98
43) Isopropyl Acetate	8.18	43	32640	2.865	ug/l #	79
95) Naphthalene	15.14	128	8299	2.143	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053758.D
Acq On : 13 Feb 2019 16:35
Operator : JC/SP
Sample : K1343-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-021119-I

Quant Time: Feb 14 01:13:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

