

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053761.D
 Acq On : 13 Feb 2019 17:56
 Operator : JC/SP
 Sample : K1343-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-021119-K

Quant Time: Feb 14 01:18:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	662467	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1146763	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1083264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	445827	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	370225	55.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.42%	
35) Dibromofluoromethane	7.59	113	362912	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
50) Toluene-d8	10.09	98	1402147	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
62) 4-Bromofluorobenzene	12.40	95	472707	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
Target Compounds						
16) Acetone	3.82	43	14546	8.533	ug/l	96
20) Methylene Chloride	4.55	84	36186	5.060	ug/l	91
43) Isopropyl Acetate	8.17	43	31673	2.841	ug/l #	80
95) Naphthalene	15.14	128	18307	2.639	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053761.D
Acq On : 13 Feb 2019 17:56
Operator : JC/SP
Sample : K1343-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-021119-K

Quant Time: Feb 14 01:18:28 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

