

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053763.D
 Acq On : 13 Feb 2019 18:49
 Operator : JC/SP
 Sample : K1343-26
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-021119-M

Quant Time: Feb 14 01:21:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	668358	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1153206	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1040055	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	417614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	363858	54.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.54%	
35) Dibromofluoromethane	7.59	113	360376	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	10.09	98	1459846	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
62) 4-Bromofluorobenzene	12.40	95	446244	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
Target Compounds						
16) Acetone	3.82	43	13367	7.772	ug/l	94
20) Methylene Chloride	4.55	84	2200303	304.946	ug/l	99
43) Isopropyl Acetate	8.17	43	27374	2.441	ug/l #	83
95) Naphthalene	15.13	128	18036	2.685	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053763.D
Acq On : 13 Feb 2019 18:49
Operator : JC/SP
Sample : K1343-26
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-021119-M

Quant Time: Feb 14 01:21:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

