

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053746.D
 Acq On : 13 Feb 2019 10:47
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 00:32:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.515	0.535	-3.9	94	0.00
3 P	Chloromethane	0.594	0.562	5.4	93	0.00
4 C	Vinyl Chloride	0.576	0.593	-3.0#	100	0.00
5 T	Bromomethane	0.377	0.401	-6.4	107	0.00
6 T	Chloroethane	0.329	0.369	-12.2	111	0.00
7 T	Trichlorofluoromethane	0.781	0.820	-5.0	104	0.00
8 T	Diethyl Ether	0.285	0.287	-0.7	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.509	-5.4	106	0.00
10 T	Methyl Iodide	0.759	0.720	5.1	94	0.00
11 T	Tert butyl alcohol	0.030	0.031	-3.3	107	0.00
12 CM	1,1-Dichloroethene	0.463	0.456	1.5#	97	0.00
13 T	Acrolein	0.016	0.016	0.0	119	0.00
14 T	Allyl chloride	0.737	0.717	2.7	95	0.00
15 T	Acrylonitrile	0.160	0.173	-8.1	105	0.00
16 T	Acetone	0.129	0.165	-27.9#	135	0.00
17 T	Carbon Disulfide	1.442	1.284	11.0	89	0.00
18 T	Methyl Acetate	0.381	0.430	-12.9	117	0.00
19 T	Methyl tert-butyl Ether	1.174	1.250	-6.5	103	0.00
20 T	Methylene Chloride	0.540	0.525	2.8	101	0.00
21 T	trans-1,2-Dichloroethene	0.500	0.491	1.8	98	0.00
22 T	Diisopropyl ether	1.366	1.462	-7.0	103	0.00
23 T	Vinyl Acetate	0.945	0.984	-4.1	98	0.00
24 P	1,1-Dichloroethane	0.911	0.914	-0.3	101	0.00
25 T	2-Butanone	0.183	0.204	-11.5	109	0.00
26 T	2,2-Dichloropropane	0.696	0.705	-1.3	101	0.00
27 T	cis-1,2-Dichloroethene	0.555	0.569	-2.5	100	0.00
28 T	Bromochloromethane	0.395	0.396	-0.3	100	0.00
29 T	Tetrahydrofuran	0.118	0.127	-7.6	103	0.00
30 C	Chloroform	0.903	0.920	-1.9#	104	0.00
31 T	Cyclohexane	0.949	0.808	14.9	96	0.00
32 T	1,1,1-Trichloroethane	0.752	0.781	-3.9	103	0.00
33 S	1,2-Dichloroethane-d4	0.502	0.496	1.2	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.316	0.336	-6.3	99	0.00
36 T	1,1-Dichloropropene	0.451	0.498	-10.4	103	0.00
37 T	Ethyl Acetate	0.252	0.287	-13.9	98	0.00
38 T	Carbon Tetrachloride	0.454	0.495	-9.0	102	0.00
39 T	Methylcyclohexane	0.535	0.591	-10.5	100	0.00
40 TM	Benzene	1.390	1.475	-6.1	99	0.00
41 T	Methacrylonitrile	0.141	0.174	-23.4	116	0.00
42 TM	1,2-Dichloroethane	0.413	0.443	-7.3	102	0.00
43 T	Isopropyl Acetate	0.486	0.517	-6.4	102	0.00
44 TM	Trichloroethene	0.396	0.422	-6.6	103	0.00
45 C	1,2-Dichloropropane	0.356	0.392	-10.1#	102	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
 Data File : VN053746.D
 Acq On : 13 Feb 2019 10:47
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 00:32:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.232	-8.4	103	0.00
47 T	Bromodichloromethane	0.446	0.499	-11.9	106	0.00
48 T	Methyl methacrylate	0.232	0.247	-6.5	100	0.00
49 T	1,4-Dioxane	0.003	0.004	-33.3#	110	0.00
50 S	Toluene-d8	1.186	1.249	-5.3	98	0.00
51 T	4-Methyl-2-Pentanone	0.239	0.282	-18.0	105	0.00
52 CM	Toluene	0.825	0.910	-10.3#	101	0.00
53 T	t-1,3-Dichloropropene	0.446	0.503	-12.8	103	0.00
54 T	cis-1,3-Dichloropropene	0.529	0.586	-10.8	102	0.00
55 T	1,1,2-Trichloroethane	0.297	0.345	-16.2	107	0.00
56 T	Ethyl methacrylate	0.352	0.408	-15.9	103	0.00
57 T	1,3-Dichloropropane	0.508	0.562	-10.6	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.196	0.193	1.5	85	0.00
59 T	2-Hexanone	0.162	0.189	-16.7	106	0.00
60 T	Dibromochloromethane	0.335	0.391	-16.7	107	0.00
61 T	1,2-Dibromoethane	0.294	0.325	-10.5	104	0.00
62 S	4-Bromofluorobenzene	0.396	0.437	-10.4	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.429	0.454	-5.8	105	0.00
65 PM	Chlorobenzene	1.062	1.185	-11.6	111	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.439	-18.3	115	0.00
67 C	Ethyl Benzene	1.745	1.974	-13.1#	109	0.00
68 T	m/p-Xylenes	0.666	0.759	-14.0	108	0.00
69 T	o-Xylene	0.643	0.749	-16.5	110	0.00
70 T	Styrene	1.043	1.232	-18.1	111	0.00
71 P	Bromoform	0.254	0.309	-21.7	114	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.515	3.628	-3.2	108	0.00
74 T	N-amyl acetate	0.845	0.894	-5.8	109	0.00
75 P	1,1,2,2-Tetrachloroethane	0.841	0.872	-3.7	113	0.00
76 T	1,2,3-Trichloropropane	0.642	0.717	-11.7	120	0.00
77 T	Bromobenzene	0.947	0.988	-4.3	110	0.00
78 T	n-propylbenzene	4.063	4.333	-6.6	111	0.00
79 T	2-Chlorotoluene	2.411	2.499	-3.6	110	0.00
80 T	1,3,5-Trimethylbenzene	2.915	3.162	-8.5	111	0.00
81 T	trans-1,4-Dichloro-2-butene	0.228	0.233	-2.2	104	0.00
82 T	4-Chlorotoluene	2.474	2.600	-5.1	111	0.00
83 T	tert-Butylbenzene	2.576	2.739	-6.3	111	0.00
84 T	1,2,4-Trimethylbenzene	2.970	3.231	-8.8	111	0.00
85 T	sec-Butylbenzene	3.506	4.002	-14.1	118	0.00
86 T	p-Isopropyltoluene	3.066	3.536	-15.3	117	0.00
87 T	1,3-Dichlorobenzene	1.703	1.834	-7.7	115	0.00
88 T	1,4-Dichlorobenzene	1.685	1.815	-7.7	115	0.00
89 T	n-Butylbenzene	2.654	3.041	-14.6	117	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021319\
Data File : VN053746.D
Acq On : 13 Feb 2019 10:47
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 14 00:32:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.506	0.588	-16.2	124	0.00
91 T	1,2-Dichlorobenzene	1.630	1.771	-8.7	114	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.132	-10.0	114	0.00
93 T	1,2,4-Trichlorobenzene	0.935	1.013	-8.3	105	0.00
94 T	Hexachlorobutadiene	0.582	0.658	-13.1	118	0.00
95 T	Naphthalene	1.905	2.006	-5.3	96	0.00
96 T	1,2,3-Trichlorobenzene	0.868	0.941	-8.4	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6