

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021320\
 Data File : VN060088.D
 Acq On : 13 Feb 2020 12:43
 Operator : JC/MD
 Sample : L1370-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-021220

Quant Time: Feb 14 05:40:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	207505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	326672	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	308787	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	138845	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	119943	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
35) Dibromofluoromethane	7.57	113	97572	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
50) Toluene-d8	10.08	98	395185	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	12.40	95	141693	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
Target Compounds						
16) Acetone	3.79	43	5534	5.025	ug/l	93
20) Methylene Chloride	4.53	84	28761	13.089	ug/l	99
43) Isopropyl Acetate	8.14	43	29272	6.262	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021320\
Data File : VN060088.D
Acq On : 13 Feb 2020 12:43
Operator : JC/MD
Sample : L1370-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-021220

Quant Time: Feb 14 05:40:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

