

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021320\
 Data File : VN060094.D
 Acq On : 13 Feb 2020 14:59
 Operator : JC/MD
 Sample : PB126782ZHE#14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126782ZHE#14

Quant Time: Feb 14 06:06:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	189912	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	300091	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	280717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	127452	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	111938	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
35) Dibromofluoromethane	7.57	113	90999	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
50) Toluene-d8	10.08	98	364573	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.40	95	129672	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
Target Compounds						
20) Methylene Chloride	4.52	84	31851	15.838	ug/l	98
43) Isopropyl Acetate	8.14	43	32231	7.506	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021320\
Data File : VN060094.D
Acq On : 13 Feb 2020 14:59
Operator : JC/MD
Sample : PB126782ZHE#14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126782ZHE#14

Quant Time: Feb 14 06:06:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

