

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
 Data File : VN076631.D
 Acq On : 13 Feb 2023 16:22
 Operator : JC\MD
 Sample : VN0213WBL01
 Misc : 5.0mL/MSVOA_N/Water
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0213WBL01

Quant Time: Feb 14 05:23:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	326493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	573745	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	508047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	202045	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	222696	49.396	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.800%
35) Dibromofluoromethane	8.172	113	177201	47.161	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.320%
50) Toluene-d8	10.572	98	693126	49.396	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.800%
62) 4-Bromofluorobenzene	12.854	95	203698	43.830	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
Data File : VN076631.D
Acq On : 13 Feb 2023 16:22
Operator : JC\MD
Sample : VN0213WBL01
Misc : 5.0mL/MSVOA_N/Water
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0213WBL01

Quant Time: Feb 14 05:23:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

