

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
 Data File : VN076637.D
 Acq On : 13 Feb 2023 19:08
 Operator : JC\MD
 Sample : 01532-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-021023

Quant Time: Feb 14 05:26:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	331819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	596149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	532962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.789	152	213957	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	232443	50.730	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.460%
35) Dibromofluoromethane	8.172	113	190750	48.859	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.720%
50) Toluene-d8	10.572	98	722088	49.526	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.060%
62) 4-Bromofluorobenzene	12.854	95	221527	45.875	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.740%
Target Compounds						
16) Acetone	4.454	43	17836	8.715	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
Data File : VN076637.D
Acq On : 13 Feb 2023 19:08
Operator : JC\MD
Sample : 01532-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-021023

Quant Time: Feb 14 05:26:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

