

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
 Data File : VN076646.D
 Acq On : 13 Feb 2023 22:44
 Operator : JC\MD
 Sample : 01510-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-51B-020823

Quant Time: Feb 14 05:32:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	353079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	619568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	580074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	227710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	231463	47.475	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.940%
35) Dibromofluoromethane	8.172	113	186536	45.974	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.940%
50) Toluene-d8	10.572	98	760000	50.156	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.320%
62) 4-Bromofluorobenzene	12.848	95	237466	47.317	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.640%
Target Compounds						
					Qvalue	
16) Acetone	4.437	43	97534	44.785	ug/l	91
20) Methylene Chloride	5.290	84	26189	5.738	ug/l	93
37) Ethyl Acetate	7.566	43	59848	9.703	ug/l	97
43) Isopropyl Acetate	8.689	43	306352	30.523	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
Data File : VN076646.D
Acq On : 13 Feb 2023 22:44
Operator : JC\MD
Sample : 01510-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-51B-020823

Quant Time: Feb 14 05:32:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

