

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
 Data File : VN076647.D
 Acq On : 13 Feb 2023 23:08
 Operator : JC\MD
 Sample : 01510-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-55-020823

Quant Time: Feb 14 05:33:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	374642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	670730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	621931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	250254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	251719	48.658	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.320%
35) Dibromofluoromethane	8.172	113	203375	46.301	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.600%
50) Toluene-d8	10.571	98	815164	49.693	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.380%
62) 4-Bromofluorobenzene	12.854	95	258787	47.632	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.260%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	90043	38.966	ug/l	92
20) Methylene Chloride	5.301	84	30718	6.343	ug/l	86
37) Ethyl Acetate	7.572	43	59029	8.840	ug/l	97
43) Isopropyl Acetate	8.695	43	325919	29.996	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
Data File : VN076647.D
Acq On : 13 Feb 2023 23:08
Operator : JC\MD
Sample : 01510-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-55-020823

Quant Time: Feb 14 05:33:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

