

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
 Data File : VN076648.D
 Acq On : 13 Feb 2023 23:32
 Operator : JC\MD
 Sample : 01510-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-59-020823

Quant Time: Feb 14 05:35:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	340887	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	626678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	575479	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	237793	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	236228	50.185	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.360%
35) Dibromofluoromethane	8.172	113	186798	45.516	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.040%
50) Toluene-d8	10.572	98	757907	49.450	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.900%
62) 4-Bromofluorobenzene	12.848	95	236791	46.647	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.300%
Target Compounds						
					Qvalue	
16) Acetone	4.454	43	101890	48.458	ug/l #	80
20) Methylene Chloride	5.284	84	30718	6.971	ug/l #	90
37) Ethyl Acetate	7.572	43	54724	8.772	ug/l	98
43) Isopropyl Acetate	8.695	43	390827	38.498	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
Data File : VN076648.D
Acq On : 13 Feb 2023 23:32
Operator : JC\MD
Sample : 01510-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-59-020823

Quant Time: Feb 14 05:35:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

