

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
 Data File : VN076649.D
 Acq On : 13 Feb 2023 23:56
 Operator : JC\MD
 Sample : 01489-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BARN-EAST-END-DRAIN

Quant Time: Feb 14 05:36:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	339308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	623476	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	588579	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	231879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	229474	48.977	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.960%
35) Dibromofluoromethane	8.172	113	188746	46.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.460%
50) Toluene-d8	10.572	98	772159	50.639	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.280%
62) 4-Bromofluorobenzene	12.854	95	240194	47.560	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.120%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	131203	62.690	ug/l	94
20) Methylene Chloride	5.289	84	32800	7.478	ug/l	89
25) 2-Butanone	7.501	43	31633	10.247	ug/l	96
37) Ethyl Acetate	7.572	43	49624	7.995	ug/l	97
43) Isopropyl Acetate	8.689	43	411022	40.695	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
Data File : VN076649.D
Acq On : 13 Feb 2023 23:56
Operator : JC\MD
Sample : 01489-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BARN-EAST-END-DRAIN

Quant Time: Feb 14 05:36:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

