

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
 Data File : VN076630.D
 Acq On : 13 Feb 2023 15:48
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 05:22:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.531	0.535	-0.8	92	0.00
3 P	Chloromethane	0.693	0.617	11.0	95	0.00
4 C	Vinyl Chloride	0.682	0.665	2.5#	96	0.00
5 T	Bromomethane	0.413	0.418	-1.2	111	0.00
6 T	Chloroethane	0.431	0.421	2.3	94	0.00
7 T	Trichlorofluoromethane	0.928	0.873	5.9	95	0.00
8 T	Diethyl Ether	0.357	0.357	0.0	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.539	1.1	102	0.00
10 T	Methyl Iodide	0.673	0.669	0.6	84	0.00
11 T	Tert butyl alcohol	0.091	0.092	-1.1	99	0.00
12 CM	1,1-Dichloroethene	0.516	0.489	5.2#	96	0.00
13 T	Acrolein	0.137	0.136	0.7	97	0.00
14 T	Allyl chloride	0.894	0.877	1.9	97	0.00
15 T	Acrylonitrile	0.323	0.325	-0.6	98	0.00
16 T	Acetone	0.308	0.281	8.8	100	0.00
17 T	Carbon Disulfide	1.363	1.321	3.1	96	0.00
18 T	Methyl Acetate	1.007	0.990	1.7	97	0.00
19 T	Methyl tert-butyl Ether	1.740	1.798	-3.3	98	0.00
20 T	Methylene Chloride	0.646	0.593	8.2	94	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.548	5.7	97	0.00
22 T	Diisopropyl ether	2.097	2.181	-4.0	98	0.00
23 T	Vinyl Acetate	1.241	1.484	-19.6	103	0.00
24 P	1,1-Dichloroethane	1.149	1.105	3.8	98	0.00
25 T	2-Butanone	0.455	0.466	-2.4	101	0.00
26 T	2,2-Dichloropropane	0.829	0.871	-5.1	103	0.00
27 T	cis-1,2-Dichloroethene	0.689	0.665	3.5	96	0.00
28 T	Bromochloromethane	0.547	0.531	2.9	95	0.00
29 T	Tetrahydrofuran	0.297	0.308	-3.7	99	0.00
30 C	Chloroform	1.148	1.122	2.3#	98	0.00
31 T	Cyclohexane	1.083	1.035	4.4	97	0.00
32 T	1,1,1-Trichloroethane	0.966	0.953	1.3	96	0.00
33 S	1,2-Dichloroethane-d4	0.690	0.659	4.5	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.327	0.306	6.4	95	0.00
36 T	1,1-Dichloropropene	0.473	0.482	-1.9	99	0.00
37 T	Ethyl Acetate	0.498	0.537	-7.8	101	0.00
38 T	Carbon Tetrachloride	0.477	0.479	-0.4	98	0.00
39 T	Methylcyclohexane	0.538	0.603	-12.1	102	0.00
40 TM	Benzene	1.453	1.479	-1.8	98	0.00
41 T	Methacrylonitrile	0.288	0.307	-6.6	98	0.00
42 TM	1,2-Dichloroethane	0.507	0.490	3.4	95	0.00
43 T	Isopropyl Acetate	0.810	0.886	-9.4	99	0.00
44 TM	Trichloroethene	0.343	0.357	-4.1	98	0.00
45 C	1,2-Dichloropropane	0.385	0.389	-1.0#	98	0.00
46 T	Dibromomethane	0.246	0.255	-3.7	100	0.00
47 T	Bromodichloromethane	0.493	0.513	-4.1	100	0.00
48 T	Methyl methacrylate	0.384	0.417	-8.6	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
 Data File : VN076630.D
 Acq On : 13 Feb 2023 15:48
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 14 05:22:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	109	0.00
50 S	Toluene-d8	1.223	1.211	1.0	96	0.00
51 T	4-Methyl-2-Pentanone	0.515	0.557	-8.2	101	0.00
52 CM	Toluene	0.877	0.946	-7.9#	100	0.00
53 T	t-1,3-Dichloropropene	0.476	0.537	-12.8	100	0.00
54 T	cis-1,3-Dichloropropene	0.549	0.588	-7.1	98	0.00
55 T	1,1,2-Trichloroethane	0.357	0.369	-3.4	100	0.00
56 T	Ethyl methacrylate	0.502	0.571	-13.7	98	0.00
57 T	1,3-Dichloropropane	0.613	0.630	-2.8	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.230	0.266	-15.7	99	0.00
59 T	2-Hexanone	0.372	0.407	-9.4	99	0.00
60 T	Dibromochloromethane	0.376	0.391	-4.0	97	0.00
61 T	1,2-Dibromoethane	0.357	0.371	-3.9	100	0.00
62 S	4-Bromofluorobenzene	0.405	0.426	-5.2	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.315	0.324	-2.9	100	0.00
65 PM	Chlorobenzene	1.060	1.055	0.5	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.389	-1.0	95	0.00
67 C	Ethyl Benzene	1.805	1.918	-6.3#	100	0.00
68 T	m/p-Xylenes	0.696	0.747	-7.3	98	0.00
69 T	o-Xylene	0.679	0.742	-9.3	100	0.00
70 T	Styrene	1.097	1.238	-12.9	100	0.00
71 P	Bromoform	0.282	0.304	-7.8	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.665	3.810	-4.0	99	0.00
74 T	N-amyl acetate	1.652	1.701	-3.0	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.327	1.189	10.4	97	0.00
76 T	1,2,3-Trichloropropane	1.011	0.901	10.9	90	0.00
77 T	Bromobenzene	0.914	0.881	3.6	98	0.00
78 T	n-propylbenzene	4.303	4.586	-6.6	100	0.00
79 T	2-Chlorotoluene	2.635	2.617	0.7	99	0.00
80 T	1,3,5-Trimethylbenzene	3.029	3.242	-7.0	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.347	0.353	-1.7	95	0.00
82 T	4-Chlorotoluene	2.635	2.617	0.7	99	0.00
83 T	tert-Butylbenzene	2.740	2.842	-3.7	97	0.00
84 T	1,2,4-Trimethylbenzene	3.019	3.260	-8.0	101	0.00
85 T	sec-Butylbenzene	3.825	4.170	-9.0	100	0.00
86 T	p-Isopropyltoluene	3.055	3.435	-12.4	99	0.00
87 T	1,3-Dichlorobenzene	1.728	1.714	0.8	98	0.00
88 T	1,4-Dichlorobenzene	1.721	1.716	0.3	99	0.00
89 T	n-Butylbenzene	2.657	3.049	-14.8	101	0.00
90 T	Hexachloroethane	0.640	0.617	3.6	97	0.00
91 T	1,2-Dichlorobenzene	1.772	1.729	2.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.237	7.8	100	0.00
93 T	1,2,4-Trichlorobenzene	0.884	1.039	-17.5	105	0.00
94 T	Hexachlorobutadiene	0.433	0.460	-6.2	102	0.00
95 T	Naphthalene	2.682	3.338	-24.5	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021323\
Data File : VN076630.D
Acq On : 13 Feb 2023 15:48
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 14 05:22:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.856	1.025	-19.7	105	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6