

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021419\
 Data File : VN053781.D
 Acq On : 14 Feb 2019 16:02
 Operator : JC/SP
 Sample : PB117097ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117097ZHE#01

Quant Time: Feb 15 04:28:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	757840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1262841	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1189715	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	462770	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	401609	52.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.66%	
35) Dibromofluoromethane	7.59	113	398801	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
50) Toluene-d8	10.09	98	1534305	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	12.40	95	508393	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
Target Compounds						
16) Acetone	3.82	43	14536	7.454	ug/l	100
20) Methylene Chloride	4.55	84	17501	2.139	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021419\
Data File : VN053781.D
Acq On : 14 Feb 2019 16:02
Operator : JC/SP
Sample : PB117097ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117097ZHE#01

Quant Time: Feb 15 04:28:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

