

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021419\
 Data File : VN053787.D
 Acq On : 14 Feb 2019 18:42
 Operator : JC/SP
 Sample : PB117097ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117097ZHE#07

Quant Time: Feb 15 04:32:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	733258	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1230963	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1157995	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	463286	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	390462	53.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.16%	
35) Dibromofluoromethane	7.59	113	389220	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
50) Toluene-d8	10.09	98	1505565	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
62) 4-Bromofluorobenzene	12.40	95	496874	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
Target Compounds						
16) Acetone	3.82	43	15596	8.266	ug/l #	80
20) Methylene Chloride	4.56	84	16614	2.099	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021419\
Data File : VN053787.D
Acq On : 14 Feb 2019 18:42
Operator : JC/SP
Sample : PB117097ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117097ZHE#07

Quant Time: Feb 15 04:32:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

