

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021420\
 Data File : VN060099.D
 Acq On : 14 Feb 2020 10:10
 Operator : JC/MD
 Sample : VN0214MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0214MBL01

Quant Time: Feb 15 02:50:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	162654	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	257694	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	222923	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	80638	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	98601	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	
35) Dibromofluoromethane	7.56	113	78227	49.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.76%	
50) Toluene-d8	10.08	98	306376	49.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.24%	
62) 4-Bromofluorobenzene	12.40	95	96777	42.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.44%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021420\
Data File : VN060099.D
Acq On : 14 Feb 2020 10:10
Operator : JC/MD
Sample : VN0214MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0214MBL01

Quant Time: Feb 15 02:50:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

