

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
 Data File : VN076658.D
 Acq On : 14 Feb 2023 13:41
 Operator : JC\MD
 Sample : PB150840TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150840TB

Quant Time: Feb 14 15:23:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	321086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	609183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	578129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	231496	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	225542	50.870	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.740%		
35) Dibromofluoromethane	8.172	113	184611	46.275	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.560%		
50) Toluene-d8	10.572	98	763466	51.244	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.480%		
62) 4-Bromofluorobenzene	12.848	95	239490	48.534	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 97.060%		
Target Compounds						
						Qvalue
16) Acetone	4.454	43	40194	20.295	ug/l #	83
20) Methylene Chloride	5.301	84	13843	3.335	ug/l #	86
25) 2-Butanone	7.489	43	18390	6.295	ug/l	93
37) Ethyl Acetate	7.572	43	60966	10.053	ug/l	97
43) Isopropyl Acetate	8.695	43	388287	39.346	ug/l #	81
52) Toluene	10.630	92	12025	1.125	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
Data File : VN076658.D
Acq On : 14 Feb 2023 13:41
Operator : JC\MD
Sample : PB150840TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150840TB

Quant Time: Feb 14 15:23:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

