

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
 Data File : VN076659.D
 Acq On : 14 Feb 2023 14:05
 Operator : JC\MD
 Sample : 01541-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ASH

Quant Time: Feb 14 15:23:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	330754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	615138	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	582596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	229092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	230768	50.527	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.060%
35) Dibromofluoromethane	8.177	113	183676	45.595	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.180%
50) Toluene-d8	10.571	98	761763	50.634	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.260%
62) 4-Bromofluorobenzene	12.854	95	241963	48.560	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.120%
Target Compounds						
						Qvalue
16) Acetone	4.448	43	45591	22.347	ug/l	94
20) Methylene Chloride	5.289	84	20279	4.743	ug/l	96
25) 2-Butanone	7.489	43	18300	6.081	ug/l	95
37) Ethyl Acetate	7.572	43	50339	8.220	ug/l	97
43) Isopropyl Acetate	8.689	43	366734	36.803	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
Data File : VN076659.D
Acq On : 14 Feb 2023 14:05
Operator : JC\MD
Sample : 01541-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ASH

Quant Time: Feb 14 15:23:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

