

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
 Data File : VN076662.D
 Acq On : 14 Feb 2023 15:17
 Operator : JC\MD
 Sample : 01519-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW10-EB-20230209

Quant Time: Feb 14 15:35:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	312294	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	585894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	508919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	204875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	228165	52.910	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.820%
35) Dibromofluoromethane	8.172	113	177419	46.240	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.480%
50) Toluene-d8	10.572	98	698673	48.759	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.520%
62) 4-Bromofluorobenzene	12.854	95	210569	44.369	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.740%
Target Compounds						
16) Acetone	4.454	43	15716	8.159	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
Data File : VN076662.D
Acq On : 14 Feb 2023 15:17
Operator : JC\MD
Sample : 01519-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW10-EB-20230209

Quant Time: Feb 14 15:35:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

