

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
 Data File : VN076670.D
 Acq On : 14 Feb 2023 19:00
 Operator : JC\MD
 Sample : 01549-08 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1696

Quant Time: Feb 15 03:17:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	326490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	593671	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	523483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	212606	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	231116	51.264	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.520%
35) Dibromofluoromethane	8.172	113	177605	45.682	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.360%
50) Toluene-d8	10.572	98	718194	49.465	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.920%
62) 4-Bromofluorobenzene	12.848	95	226166	47.031	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
Data File : VN076670.D
Acq On : 14 Feb 2023 19:00
Operator : JC\MD
Sample : 01549-08 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1696

Quant Time: Feb 15 03:17:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

