

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
 Data File : VN076660.D
 Acq On : 14 Feb 2023 14:29
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 14 14:51:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	355837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	650598	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	553040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	226644	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	246840	50.236	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.480%
35) Dibromofluoromethane	8.172	113	196171	46.042	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.080%
50) Toluene-d8	10.572	98	775486	48.737	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.480%
62) 4-Bromofluorobenzene	12.848	95	229763	43.598	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.200%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021423\
Data File : VN076660.D
Acq On : 14 Feb 2023 14:29
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 14 14:51:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.231 min Scan# 1070
Delta R.T. -0.000 min
Lab File: VN076660.D
Acq: 14 Feb 2023 14:29




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 355837
Ion Ratio Lower Upper
168 100
99 61.0 44.8 67.2



#33
1,2-Dichloroethane-d4
Concen: 50.236 ug/l
RT: 8.589 min Scan# 1131
Delta R.T. 0.006 min
Lab File: VN076660.D
Acq: 14 Feb 2023 14:29

Tgt Ion: 65 Resp: 246840
Ion Ratio Lower Upper
65 100
67 53.9 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.107 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN076660.D

Acq: 14 Feb 2023 14:29

Instrument :

MSVOA_N

ClientSampleId :

IBLK





Tgt Ion:114 Resp: 650598

Ion Ratio Lower Upper

114 100

63 22.2 0.0 40.0

88 14.6 0.0 29.6



#35

Dibromofluoromethane

Concen: 46.042 ug/l

RT: 8.172 min Scan# 1060

Delta R.T. -0.000 min

Lab File: VN076660.D

Acq: 14 Feb 2023 14:29




Tgt Ion:113 Resp: 196171


Ion Ratio Lower Upper

113 100

111 102.6 83.1 124.7

192 18.0 15.6 23.4









#50
Toluene-d8
Concen: 48.737 ug/l
RT: 10.572 min Scan# 1468
Delta R.T. 0.006 min
Lab File: VN076660.D
Acq: 14 Feb 2023 14:29




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 775486
Ion Ratio Lower Upper
98 100
100 64.5 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 43.598 ug/l
RT: 12.848 min Scan# 1855
Delta R.T. -0.000 min
Lab File: VN076660.D
Acq: 14 Feb 2023 14:29

Tgt Ion: 95 Resp: 229763
Ion Ratio Lower Upper
95 100
174 78.1 0.0 169.2
176 74.8 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.872 min Scan# 1689
Delta R.T. 0.006 min
Lab File: VN076660.D
Acq: 14 Feb 2023 14:29

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 553040
Ion Ratio Lower Upper
117 100
82 54.0 43.4 65.2
119 31.1 26.1 39.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.795 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN076660.D
Acq: 14 Feb 2023 14:29

Tgt Ion:152 Resp: 226644
Ion Ratio Lower Upper
152 100
115 58.0 29.3 87.9
150 156.2 0.0 349.8

