

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021424\
 Data File : VN081032.D
 Acq On : 14 Feb 2024 18:07
 Operator : JC\MD
 Sample : P1459-03 20X
 Misc : 1.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WAL-OILY-DEBRIS

Quant Time: Feb 15 08:04:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 07:59:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	562248	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	991786	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	910739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	349543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	332638	48.099	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.200%
35) Dibromofluoromethane	8.165	113	273768	48.144	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	10.571	98	989699	49.350	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.700%
62) 4-Bromofluorobenzene	12.853	95	370507	54.322	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.640%
Target Compounds						
					Qvalue	
52) Toluene	10.629	92	40043	2.285	ug/l	99
67) Ethyl Benzene	11.965	91	82747	2.298	ug/l	96
68) m/p-Xylenes	12.065	106	96500	7.268	ug/l	99
69) o-Xylene	12.400	106	46551	3.615	ug/l	90
78) n-propylbenzene	13.035	91	49631	1.232	ug/l	85
80) 1,3,5-Trimethylbenzene	13.176	105	101957	3.721	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	316040	11.971	ug/l	99
85) sec-Butylbenzene	13.617	105	40848	1.269	ug/l #	75
86) p-Isopropyltoluene	13.735	119	36237	1.400	ug/l	90
89) n-Butylbenzene	14.059	91	31023	1.309	ug/l #	86
95) Naphthalene	15.641	128	161319	6.419	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021424\
Data File : VN081032.D
Acq On : 14 Feb 2024 18:07
Operator : JC\MD
Sample : P1459-03 20X
Misc : 1.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WAL-OILY-DEBRIS

Quant Time: Feb 15 08:04:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 07:59:19 2024
Response via : Initial Calibration

