



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021424\
Data File : VN081017.D
Acq On : 14 Feb 2024 11:15
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
Title : SW846 8260
Last Update : Thu Feb 15 07:59:19 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1842

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.8	27688	PASS
75	95	30	60	55.2	73440	PASS
95	95	100	100	100.0	133155	PASS
96	95	5	9	6.8	9110	PASS
173	174	0.00	2	0.9	816	PASS
174	95	50	100	71.3	94901	PASS
175	174	5	9	7.6	7174	PASS
176	174	95	101	95.5	90640	PASS
177	176	5	9	6.4	5773	PASS

Average of 12.841 to 12.853 min.: VN081017.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.80	291	46.00	107	55.10	177	64.90	51
36.00	1051	46.30	94	56.05	2369	66.95	72
37.05	6923	47.05	1733	56.90	4076	68.00	1413
37.95	6786	47.85	1032	57.90	142	69.00	1419
39.00	2983	49.00	4951	58.30	80	70.00	1359
39.95	364	50.00	27688	60.00	1554	71.90	884
41.00	128	51.00	8947	61.00	7305	72.95	7362
42.85	479	51.95	505	61.95	7272	74.00	25669
43.95	1606	52.90	77	62.80	1301	75.00	73440
44.85	1029	53.20	184	63.00	3801	75.95	5680
45.10	145	54.80	437	64.00	625	76.95	637

Average of 12.841 to 12.853 min.: VN081017.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
77.70	88	88.10	1028	104.80	167	115.95	666
77.90	388	90.85	991	105.05	160	116.95	1063
78.85	5187	91.80	1002	105.80	739	117.65	530
79.90	1399	91.95	3146	106.85	145	118.00	78
80.95	6303	92.95	6225	109.65	151	118.85	820
81.95	1477	93.95	16951	110.00	156	126.60	116
82.75	250	95.00	133155	110.95	211	127.50	139
83.00	121	96.00	9110	111.70	210	127.80	381
85.65	160	96.95	242	112.75	156	128.00	192
86.95	5167	97.20	231	113.00	67	128.75	242
87.85	4736	103.80	893	114.90	73	129.60	175

Average of 12.841 to 12.853 min.: VN081017.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
129.90	320	145.00	66	156.95	234	177.90	113
130.10	149	145.60	121	158.90	111	206.75	104
130.90	303	145.90	163	160.80	51	280.90	130
134.60	79	146.85	132	171.30	99		
134.85	362	147.85	382	171.80	146		
136.85	413	148.85	197	172.10	101		
140.95	1407	149.60	78	172.90	816		
141.70	58	150.00	53	173.95	94901		
141.90	229	152.90	85	174.95	7174		
142.95	1529	153.80	74	175.95	90640		
143.90	59	154.85	505	176.90	5773		