

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021425\
 Data File : VN085769.D
 Acq On : 14 Feb 2025 14:18
 Operator : JC\MD
 Sample : PB166701ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166701ZHE#03

Quant Time: Feb 14 22:38:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	245828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	475235	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	456510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	164170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	193409	48.740	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.480%
35) Dibromofluoromethane	8.165	113	166406	50.473	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.940%
50) Toluene-d8	10.565	98	621535	53.059	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.120%
62) 4-Bromofluorobenzene	12.847	95	186965	46.659	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.320%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	43670	34.060	ug/l #	86
20) Methylene Chloride	5.265	84	3905	1.230	ug/l #	63
43) Isopropyl Acetate	8.688	43	206286	27.564	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021425\
Data File : VN085769.D
Acq On : 14 Feb 2025 14:18
Operator : JC\MD
Sample : PB166701ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166701ZHE#03

Quant Time: Feb 14 22:38:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

