

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021425\
 Data File : VN085770.D
 Acq On : 14 Feb 2025 14:42
 Operator : JC\MD
 Sample : PB166701ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166701ZHE#04

Quant Time: Feb 14 22:38:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	233308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	460844	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	437559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	150457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	189543	50.329	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.660%		
35) Dibromofluoromethane	8.165	113	161446	50.497	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.000%		
50) Toluene-d8	10.565	98	597898	52.635	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 105.280%		
62) 4-Bromofluorobenzene	12.847	95	171982	44.260	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 88.520%		
Target Compounds						
						Qvalue
16) Acetone	4.424	43	43276	35.564	ug/l	92
20) Methylene Chloride	5.271	84	4218	1.400	ug/l #	76
43) Isopropyl Acetate	8.688	43	212807	29.324	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021425\
Data File : VN085770.D
Acq On : 14 Feb 2025 14:42
Operator : JC\MD
Sample : PB166701ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166701ZHE#04

Quant Time: Feb 14 22:38:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

