

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021519\
 Data File : VN053797.D
 Acq On : 15 Feb 2019 13:57
 Operator : JC/SP
 Sample : K1484-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Feb 15 23:24:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	664107	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1219727	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1091112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	437829	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	378930	56.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.76%	
35) Dibromofluoromethane	7.59	113	370426	48.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.12%	
50) Toluene-d8	10.09	98	1449459	50.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.18%	
62) 4-Bromofluorobenzene	12.40	95	469455	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	

Target Compounds


						Qvalue
14) Allyl chloride	4.56	41	110706	11.306	ug/l #	27
16) Acetone	3.82	43	13364	7.820	ug/l	97
18) Methyl Acetate	4.33	43	3230	0.638	ug/l #	85
20) Methylene Chloride	4.55	84	2993171	417.487	ug/l	99
25) 2-Butanone	6.85	43	2468	1.013	ug/l	93
29) Tetrahydrofuran	7.22	42	1050	0.671	ug/l #	83
31) Cyclohexane	7.67	56	11067	0.816	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	41169	3.744	ug/l #	49
38) Carbon Tetrachloride	7.67	117	56996	5.146	ug/l #	16
39) Methylcyclohexane	9.08	83	3597	0.276	ug/l	96
43) Isopropyl Acetate	8.17	43	32243	2.719	ug/l #	80
45) 1,2-Dichloropropane	9.26	63	2041	0.235	ug/l #	43
48) Methyl methacrylate	9.18	41	9319	1.646	ug/l #	12
49) 1,4-Dioxane	9.27	88	66	39.871	ug/l #	1
51) 4-Methyl-2-Pentanone	10.02	43	7509	1.289	ug/l #	34
54) cis-1,3-Dichloropropene	9.90	75	66477	5.154	ug/l #	60
57) 1,3-Dichloropropane	10.76	76	12993	1.048	ug/l #	43
59) 2-Hexanone	10.76	43	165071	41.725	ug/l #	51
71) Bromoform	12.40	173	3199	0.576	ug/l #	1
74) N-amyl acetate	11.88	43	17478	2.361	ug/l #	51
76) 1,2,3-Trichloropropane	12.40	75	214119	38.075	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	15117	0.592	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.40	75	214119	107.063	ug/l #	15
84) 1,2,4-Trimethylbenzene	13.04	105	19447	0.748	ug/l	100
90) Hexachloroethane	13.89	117	1214	0.274	ug/l #	3
95) Naphthalene	15.13	128	8256	2.151	ug/l #	88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021519\
Data File : VN053797.D
Acq On : 15 Feb 2019 13:57
Operator : JC/SP
Sample : K1484-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Feb 15 23:24:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57


Instrument :
MSVOA_N
ClientSampled :

Tot Ion:168 Resp: 664107

Ion Ratio Lower Upper

168 100

99 48.3 39.1 58.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

250000

200000



150000

100000

50000

0

Time-->



#14

Allyl chloride

Concen: 11.306 ug/l

RT: 4.56 min Scan# 866

Delta R.T. 0.23 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57


Tgt Ion: 41 Resp: 110706

Ion Ratio Lower Upper

41 100

39 0.6 48.9 73.3#

76 0.0 29.4 44.2#



Abundance Ion 41.00 (40.70 to 41.70): VNO

Ion 39.00 (38.70 to 39.70): VNO

Ion 75.90 (75.60 to 76.60): VNO

4.56

50000

40000


30000


20000

10000

0

Time-->





#16

Acetone

Concen: 7.820 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.01 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 13364

Ion Ratio Lower Upper




43 100

58 35.4 27.1 40.7



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#18

Methyl Acetate

Concen: 0.638 ug/l

RT: 4.33 min Scan# 795

Delta R.T. 0.01 min

Lab File: VN053797.D


Acq: 15 Feb 2019 13:57

Tgt Ion: 43 Resp: 3230

Ion Ratio Lower Upper



43 100


74 17.1 19.4 29.2#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0









#20
Methvlene Chloride
Concen: 417.487 ug/l
RT: 4.55 min Scan# 865
Delta R.T. 0.00 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 84 Resp: 2993171




Ion	Ratio	Lower	Upper
84	100		
49	115.8	94.3	141.5
51	36.1	29.4	44.2
86	64.7	51.4	77.0




#25
2-Butanone
Concen: 1.013 ug/l
RT: 6.85 min Scan# 1579
Delta R.T. 0.02 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57

Tgt Ion: 43 Resp: 2468

Ion	Ratio	Lower	Upper
43	100		
72	24.0	22.0	33.0









#29
Tetrahydrofuran
Concen: 0.671 ug/l
RT: 7.22 min Scan# 1693
Delta R.T. 0.01 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 42 Resp: 1050
Ion Ratio Lower Upper
42 100
72 57.0 35.2 52.8#
71 32.7 33.0 49.4#



#31
Cyclohexane
Concen: 0.816 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. 0.01 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57

Tgt Ion: 56 Resp: 11067
Ion Ratio Lower Upper
56 100
69 190.5 26.4 39.6#
84 170.8 67.9 101.9#









#33
 1,2-Dichloroethane-d4
 Concen: 56.877 ug/l
 RT: 8.03 min Scan# 1946
 Delta R.T. 0.00 min
 Lab File: VN053797.D
 Acq: 15 Feb 2019 13:57




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 65 Resp: 378930
 Ion Ratio Lower Upper
 65 100
 67 55.7 0.0 111.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN053797.D
 Acq: 15 Feb 2019 13:57

Tgt Ion: 114 Resp: 1219727
 Ion Ratio Lower Upper
 114 100
 63 18.8 0.0 38.6
 88 14.9 0.0 30.4





#35

Dibromofluoromethane

Concen: 48.063 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:113 Resp: 370426

Ion Ratio Lower Upper




113 100

111 102.0 81.1 121.7

192 23.4 18.2 27.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 3.744 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.13 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57


Tgt Ion: 75 Resp: 41169

Ion Ratio Lower Upper



75 100


110 9.1 18.1 54.3#

77 0.0 25.1 37.7#



Abundance Ion 74.90 (74.60 to 75.60): VNO
Ion 109.90 (109.60 to 110.60): V
Ion 76.90 (76.60 to 77.60): VNO









#38
Carbon Tetrachloride
Concen: 5.146 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. -0.11 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:117 Resp: 56996
Ion Ratio Lower Upper
117 100
119 5.0 77.2 115.8#
121 0.0 24.6 37.0#




Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V






#39
Methylcyclohexane
Concen: 0.276 ug/l
RT: 9.08 min Scan# 2273
Delta R.T. 0.00 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57

Tgt Ion: 83 Resp: 3597
Ion Ratio Lower Upper
83 100
55 77.5 60.1 90.1
98 49.0 36.6 55.0



Abundance Ion 83.00 (82.70 to 83.70): VNO
Ion 55.00 (54.70 to 55.70): VNO
Ion 98.00 (97.70 to 98.70): VNO





#43

Isopropyl Acetate

Concen: 2.719 ug/l

RT: 8.17 min Scan# 1990

Delta R.T. 0.01 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N

ClientSampleId :





Tot Ion: 43 Resp: 32243

Ion Ratio Lower Upper

43 100

61 16.9 23.9 35.9#

87 9.8 11.0 16.6#



#45

1,2-Dichloropropane

Concen: 0.235 ug/l

RT: 9.26 min Scan# 2330

Delta R.T. 0.14 min

Lab File: VN053797.D




Acq: 15 Feb 2019 13:57


Tgt Ion: 63 Resp: 2041

Ion Ratio Lower Upper

63 100

65 0.0 25.0 37.4#









#48
Methyl methacrylate
Concen: 1.646 ug/l
RT: 9.18 min Scan# 2304
Delta R.T. -0.02 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 41 Resp: 9319
Ion Ratio Lower Upper
41 100
69 0.0 73.8 110.6#
39 0.0 40.0 60.0#



#49
1,4-Dioxane
Concen: 39.871 ug/l
RT: 9.27 min Scan# 2332
Delta R.T. 0.07 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57

Tgt Ion: 88 Resp: 66
Ion Ratio Lower Upper
88 100
43 2219916.7 27.8 41.6#
58 11395.5 56.5 84.7#





#50

Toluene-d8

Concen: 50.087 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 98 Resp: 1449459

Ion Ratio Lower Upper




98 100

100 64.3 51.4 77.0



Abundance Ion 98.00 (97.70 to 98.70): VN053561.D (-2570) (-)

Ion 100.00 (99.70 to 100.70): VN053561.D (-2570) (-)



#51

4-Methyl-2-Pentanone

Concen: 1.289 ug/l

RT: 10.02 min Scan# 2566

Delta R.T. 0.04 min

Lab File: VN053797.D


Acq: 15 Feb 2019 13:57

Tgt Ion: 43 Resp: 7509

Ion Ratio Lower Upper



43 100


58 0.0 32.8 49.2#



Abundance Ion 43.00 (42.70 to 43.70): VN053561.D (-2537) (-)

Ion 58.00 (57.70 to 58.70): VN053561.D (-2537) (-)





#54

cis-1,3-Dichloropropene

Concen: 5.154 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.07 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 75 Resp: 66477

Ion Ratio Lower Upper

75 100

77 0.6 25.8 38.6#




39 24.9 33.4 50.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0



#57

1,3-Dichloropropane

Concen: 1.048 ug/l

RT: 10.76 min Scan# 2796

Delta R.T. 0.05 min

Lab File: VN053797.D


Acq: 15 Feb 2019 13:57

Tgt Ion: 76 Resp: 12993

Ion Ratio Lower Upper



76 100


78 0.0 25.6 38.4#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0









#59
2-Hexanone
Concen: 41.725 ug/l
RT: 10.76 min Scan# 2796
Delta R.T. 0.01 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 165071
Ion Ratio Lower Upper
43 100
58 21.0 28.6 85.8#



#62
4-Bromofluorobenzene
Concen: 48.641 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.00 min
Lab File: VN053797.D
Acq: 15 Feb 2019 13:57

Tgt Ion: 95 Resp: 469455
Ion Ratio Lower Upper
95 100
174 93.8 0.0 183.2
176 90.7 0.0 177.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. 0.00 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:117 Resp: 1091112

Ion Ratio Lower Upper

117 100

82 51.8 42.9 64.3




119 32.3 26.0 39.0



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): V



#71

Bromoform

Concen: 0.576 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.27 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57


Tgt Ion:173 Resp: 3199

Ion Ratio Lower Upper

173 100

175 858.3 24.6 74.0#



254 0.0 0.2 0.2#




Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:152 Resp: 437829

Ion Ratio Lower Upper




152 100

115 55.1 27.8 83.4

150 156.1 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 2.361 ug/l

RT: 11.88 min Scan# 3142

Delta R.T. -0.19 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Tgt Ion: 43 Resp: 17478


Ion Ratio Lower Upper

43 100



70 2.0 35.4 53.2#


55 4.6 20.9 31.3#

61 12.2 20.5 30.7#



Abundance Ion 43.00 (42.70 to 43.70): V
Ion 70.00 (69.70 to 70.70): V
Ion 55.00 (54.70 to 55.70): V
Ion 61.00 (60.70 to 61.70): V





#76

1,2,3-Trichloropropane

Concen: 38.075 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. -0.15 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 75 Resp: 214119

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#



#80

1,3,5-Trimethylbenzene

Concen: 0.592 ug/l

RT: 12.73 min Scan# 3409

Delta R.T. 0.00 min

Lab File: VN053797.D




Acq: 15 Feb 2019 13:57


Tgt Ion: 105 Resp: 15117

Ion Ratio Lower Upper

105 100

120 49.9 24.9 74.6





#81

trans-1,4-Dichloro-2-butene

Concen: 107.063 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.10 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 75 Resp: 214119

Ion Ratio Lower Upper

75 100

53 0.0 69.8 104.6#




89 0.0 35.6 53.4#



Abundance Ion 74.90 (74.60 to 75.60): VN053797.D (-3181) (-)

Ion 53.00 (52.70 to 53.70): VN053797.D (-3181) (-)

Ion 88.90 (88.60 to 89.60): VN053797.D (-3181) (-)



#84

1,2,4-Trimethylbenzene

Concen: 0.748 ug/l

RT: 13.04 min Scan# 3504

Delta R.T. -0.00 min

Lab File: VN053797.D


Acq: 15 Feb 2019 13:57

Tgt Ion: 105 Resp: 19447

Ion Ratio Lower Upper



105 100


120 46.1 23.0 69.0



Abundance Ion 105.00 (104.70 to 105.70): VN053797.D (-3411) (-)

Ion 120.00 (119.70 to 120.70): VN053797.D (-3411) (-)





#90

Hexachloroethane

Concen: 0.274 ug/l

RT: 13.89 min Scan# 3768

Delta R.T. 0.01 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:117 Resp: 1214

Ion Ratio Lower Upper




117 100

201 0.0 46.8 140.4#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#95

Naphthalene

Concen: 2.151 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN053797.D

Acq: 15 Feb 2019 13:57


Tgt Ion:128 Resp: 8256

Ion Ratio Lower Upper

128 100

127 4.8 10.3 15.5#

129 11.2 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

