

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021519\
 Data File : VN053808.D
 Acq On : 15 Feb 2019 18:52
 Operator : JC/SP
 Sample : K1358-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Feb 15 23:27:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	672085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1138419	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1139362	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	427299	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	364344	54.04	ug/l	0.00
Spiked Amount			Recovery	=	108.08%	
35) Dibromofluoromethane	7.59	113	365746	50.85	ug/l	0.00
Spiked Amount			Recovery	=	101.70%	
50) Toluene-d8	10.09	98	1396098	51.69	ug/l	0.00
Spiked Amount			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	12.40	95	457856	50.83	ug/l	0.00
Spiked Amount			Recovery	=	101.66%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Acrolein	3.62	56	372	1.712	ug/l #	23
14) Allyl chloride	4.56	41	2164	0.218	ug/l #	27
16) Acetone	3.82	43	14295	8.266	ug/l	100
18) Methyl Acetate	4.34	43	3168	0.619	ug/l #	57
20) Methylene Chloride	4.56	84	68876	9.493	ug/l	94
25) 2-Butanone	6.84	43	1529	0.620	ug/l	95
29) Tetrahydrofuran	7.23	42	948	0.599	ug/l #	46
31) Cyclohexane	7.67	56	9561	0.671	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	39761	3.875	ug/l #	49
38) Carbon Tetrachloride	7.66	117	53838	5.208	ug/l #	16
43) Isopropyl Acetate	8.18	43	29008	2.621	ug/l #	81
45) 1,2-Dichloropropane	9.27	63	1659	0.205	ug/l #	43
48) Methyl methacrylate	9.18	41	9901	1.874	ug/l #	22
51) 4-Methyl-2-Pentanone	10.02	43	7517	1.382	ug/l #	34
52) Toluene	10.15	92	4324	0.230	ug/l	98
54) cis-1,3-Dichloropropene	9.91	75	60591	5.033	ug/l #	63
57) 1,3-Dichloropropane	10.76	76	12162	1.051	ug/l #	43
59) 2-Hexanone	10.76	43	164906	44.660	ug/l #	50
68) m/p-Xylenes	11.62	106	3353	0.221	ug/l	98
71) Bromoform	12.40	173	3409	0.588	ug/l #	1
76) 1,2,3-Trichloropropane	12.40	75	211945	38.617	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	5673	0.228	ug/l	93
81) trans-1,4-Dichloro-2-buten	12.40	75	211945	108.588	ug/l #	15
84) 1,2,4-Trimethylbenzene	13.04	105	7776	0.306	ug/l	98
95) Naphthalene	15.13	128	3312	1.907	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021519\
Data File : VN053808.D
Acq On : 15 Feb 2019 18:52
Operator : JC/SP
Sample : K1358-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Feb 15 23:27:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52


Instrument :
MSVOA_N
ClientSampled :

Tot Ion:168 Resp: 672085

Ion Ratio Lower Upper

168 100

99 48.1 39.1 58.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

250000

200000

150000



100000

50000

0

Time-->

7.60 7.70 7.80



#13

Acrolein

Concen: 1.712 ug/l

RT: 3.62 min Scan# 573

Delta R.T. 0.01 min

Lab File: VN053808.D


Acq: 15 Feb 2019 18:52

Tgt Ion: 56 Resp: 372

Ion Ratio Lower Upper

56 100

55 7.8 57.4 86.0#



Abundance Ion 56.00 (55.70 to 56.70): VNO

Ion 55.00 (54.70 to 55.70): VNO

3.62

300

250

200

150


100


50

0

Time-->

3.58 3.60 3.62 3.64









#14
Allvl chloride
Concen: 0.218 ug/l
RT: 4.56 min Scan# 866
Delta R.T. 0.23 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 41 Resp: 2164
Ion Ratio Lower Upper
41 100
39 0.0 48.9 73.3#
76 0.0 29.4 44.2#



#16
Acetone
Concen: 8.266 ug/l
RT: 3.82 min Scan# 638
Delta R.T. 0.01 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Tgt Ion: 43 Resp: 14295
Ion Ratio Lower Upper
43 100
58 34.1 27.1 40.7





#18

Methyl Acetate

Concen: 0.619 ug/l

RT: 4.34 min Scan# 797

Delta R.T. 0.01 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 3168

Ion Ratio Lower Upper




43 100

74 3.0 19.4 29.2#



Abundance Ion 43.00 (42.70 to 43.70): VN053808.D (-700) (-)

Ion 74.00 (73.70 to 74.70): VN053808.D (-700) (-)



#20

Methylene Chloride

Concen: 9.493 ug/l

RT: 4.56 min Scan# 866

Delta R.T. 0.01 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52

Tgt Ion: 84 Resp: 68876


Ion Ratio Lower Upper

84 100

49 110.1 94.3 141.5

51 35.4 29.4 44.2

86 59.4 51.4 77.0






Abundance Ion 83.90 (83.60 to 84.60): VN053808.D (-771) (-)

Ion 48.90 (48.60 to 49.60): VN053808.D (-771) (-)

Ion 50.90 (50.60 to 51.60): VN053808.D (-771) (-)

Ion 85.90 (85.60 to 86.60): VN053808.D (-771) (-)





#25

2-Butanone

Concen: 0.620 ug/l

RT: 6.84 min Scan# 1576

Delta R.T. 0.01 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52





Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 43 Resp: 1529

Ion Ratio Lower Upper

43 100

72 24.7 22.0 33.0



#29

Tetrahydrofuran

Concen: 0.599 ug/l

RT: 7.23 min Scan# 1696

Delta R.T. 0.02 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52




Tgt Ion: 42 Resp: 948


Ion Ratio Lower Upper

42 100

72 16.0 35.2 52.8#

71 0.0 33.0 49.4#





#31

Cvclohexane

Concen: 0.671 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.01 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52

Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 56 Resp: 9561

Ion Ratio Lower Upper




56 100

69 198.0 26.4 39.6#

84 166.5 67.9 101.9#



Abundance Ion 56.00 (55.70 to 56.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 84.00 (83.70 to 84.70): VNO



#33

1,2-Dichloroethane-d4

Concen: 54.039 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN053808.D


Acq: 15 Feb 2019 18:52

Tgt Ion: 65 Resp: 364344



Ion Ratio Lower Upper


65 100

67 55.0 0.0 111.2



Abundance Ion 64.90 (64.60 to 65.60): VNO
Ion 66.90 (66.60 to 67.60): VNO









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.59 min Scan# 2119
Delta R.T. 0.00 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:114 Resp: 1138419
Ion Ratio Lower Upper
114 100
63 18.8 0.0 38.6
88 15.1 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): V
Ion 87.90 (87.60 to 88.60): V






#35
Dibromofluoromethane
Concen: 50.845 ug/l
RT: 7.59 min Scan# 1810
Delta R.T. 0.00 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Tgt Ion:113 Resp: 365746
Ion Ratio Lower Upper
113 100
111 102.1 81.1 121.7
192 23.4 18.2 27.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 3.875 ug/l
 RT: 7.67 min Scan# 1833
 Delta R.T. -0.13 min
 Lab File: VN053808.D
 Acq: 15 Feb 2019 18:52




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion	75	Resp	39761
Ion Ratio	100	Lower	Upper
75	100		
110	8.6	18.1	54.3#
77	0.0	25.1	37.7#



#38
 Carbon Tetrachloride
 Concen: 5.208 ug/l
 RT: 7.66 min Scan# 1832
 Delta R.T. -0.11 min
 Lab File: VN053808.D
 Acq: 15 Feb 2019 18:52

Tgt Ion	117	Resp	53838
Ion Ratio <td>100</td> <td>Lower</td> <td>Upper</td>	100	Lower	Upper
117	100		
119	5.7	77.2	115.8#
121	0.0	24.6	37.0#





#43

Isopropyl Acetate

Concen: 2.621 ug/l

RT: 8.18 min Scan# 1991

Delta R.T. 0.01 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 43 Resp: 29008

Ion Ratio Lower Upper

43 100

61 16.6 23.9 35.9#




87 10.9 11.0 16.6#



Abundance Ion 43.00 (42.70 to 43.70): VN053808.D (-1991) (-)

Ion 61.00 (60.70 to 61.70): VN053808.D (-1991) (-)

Ion 87.00 (86.70 to 87.70): VN053808.D (-1991) (-)



#45

1,2-Dichloropropane

Concen: 0.205 ug/l

RT: 9.27 min Scan# 2331

Delta R.T. 0.15 min

Lab File: VN053808.D


Acq: 15 Feb 2019 18:52

Tgt Ion: 63 Resp: 1659

Ion Ratio Lower Upper



63 100


65 0.0 25.0 37.4#



Abundance Ion 62.90 (62.60 to 63.60): VN053808.D (-2331) (-)

Ion 64.90 (64.60 to 65.60): VN053808.D (-2331) (-)









#48
Methvl methacrylate
Concen: 1.874 ug/l
RT: 9.18 min Scan# 2302
Delta R.T. -0.02 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 41 Resp: 9901
Ion Ratio Lower Upper
41 100
69 0.0 73.8 110.6#
39 20.8 40.0 60.0#



#50
Toluene-d8
Concen: 51.689 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Tgt Ion: 98 Resp: 1396098
Ion Ratio Lower Upper
98 100
100 63.9 51.4 77.0





#51

4-Methyl-2-Pentanone

Concen: 1.382 ug/l

RT: 10.02 min Scan# 2566

Delta R.T. 0.04 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 7517

Ion Ratio Lower Upper




43 100

58 0.0 32.8 49.2#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#52

Toluene

Concen: 0.230 ug/l

RT: 10.15 min Scan# 2606

Delta R.T. -0.00 min

Lab File: VN053808.D


Acq: 15 Feb 2019 18:52

Tgt Ion: 92 Resp: 4324

Ion Ratio Lower Upper



92 100


91 172.1 139.9 209.9



Abundance Ion 92.00 (91.70 to 92.70): VN0

Ion 91.00 (90.70 to 91.70): VN0









#54

cis-1,3-Dichloropropene
 Concen: 5.033 ug/l
 RT: 9.91 min Scan# 2529
 Delta R.T. 0.07 min
 Lab File: VN053808.D
 Acq: 15 Feb 2019 18:52

Instrument :
 MSVOA_N
 ClientSampleId :




Tot Ion	Ratio	Lower	Upper
75	100		
77	0.5	25.8	38.6#
39	28.0	33.4	50.0#




#57

1,3-Dichloropropane
 Concen: 1.051 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.05 min
 Lab File: VN053808.D
 Acq: 15 Feb 2019 18:52

Tgt Ion	Ratio	Lower	Upper
76	100		
78	0.0	25.6	38.4#









#59
2-Hexanone
Concen: 44.660 ug/l
RT: 10.76 min Scan# 2796
Delta R.T. 0.01 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 164906
Ion Ratio Lower Upper
43 100
58 20.4 28.6 85.8#



#62
4-Bromofluorobenzene
Concen: 50.827 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.00 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Tgt Ion: 95 Resp: 457856
Ion Ratio Lower Upper
95 100
174 94.7 0.0 183.2
176 91.6 0.0 177.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2997
Delta R.T. 0.00 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Instrument :
MSVOA_N
ClientSampled :

Tot Ion:117 Resp: 1139362
Ion Ratio Lower Upper
117 100
82 52.2 42.9 64.3
119 32.1 26.0 39.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VN
Ion 118.90 (118.60 to 119.60): V






#68
m/p-Xylenes
Concen: 0.221 ug/l
RT: 11.62 min Scan# 3063
Delta R.T. 0.00 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Tgt Ion:106 Resp: 3353
Ion Ratio Lower Upper
106 100
91 205.5 161.4 242.2



Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VN









#71
Bromoform
Concen: 0.588 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.28 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Instrument :
MSVOA_N
ClientSampled :

Tot Ion:173 Resp: 3409
Ion Ratio Lower Upper
173 100
175 773.6 24.6 74.0#
254 0.0 0.2 0.2#




Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN053808.D
Acq: 15 Feb 2019 18:52

Tgt Ion:152 Resp: 427299
Ion Ratio Lower Upper
152 100
115 54.8 27.8 83.4
150 156.4 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#76

1,2,3-Trichloropropane

Concen: 38.617 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. -0.15 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 75 Resp: 211945

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#



#80

1,3,5-Trimethylbenzene

Concen: 0.228 ug/l

RT: 12.73 min Scan# 3409

Delta R.T. 0.00 min

Lab File: VN053808.D




Acq: 15 Feb 2019 18:52


Tgt Ion: 105 Resp: 5673

Ion Ratio Lower Upper

105 100

120 54.7 24.9 74.6





#81

trans-1,4-Dichloro-2-butene

Concen: 108.588 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.10 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 75 Resp: 211945

Ion Ratio Lower Upper

75 100

53 0.0 69.8 104.6#

89 0.0 35.6 53.4#






Abundance Ion 74.90 (74.60 to 75.60): VN053808.D (-3181) (-)

Ion 53.00 (52.70 to 53.70): VN053808.D (-3181) (-)

Ion 88.90 (88.60 to 89.60): VN053808.D (-3181) (-)

12.40



#84

1,2,4-Trimethylbenzene

Concen: 0.306 ug/l

RT: 13.04 min Scan# 3504

Delta R.T. 0.00 min

Lab File: VN053808.D


Acq: 15 Feb 2019 18:52

Tgt Ion: 105 Resp: 7776

Ion Ratio Lower Upper

105 100



120 47.0 23.0 69.0




Abundance Ion 105.00 (104.70 to 105.70): VN053808.D (-3411) (-)

Ion 120.00 (119.70 to 120.70): VN053808.D (-3411) (-)

13.04





#95

Naphthalene

Concen: 1.907 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN053808.D

Acq: 15 Feb 2019 18:52

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:128 Resp: 3312

Ion Ratio Lower Upper

128 100

127 7.6 10.3 15.5#

129 2.1 8.6 13.0#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

