

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021519\
 Data File : VN053810.D
 Acq On : 15 Feb 2019 19:46
 Operator : JC/SP
 Sample : PB117124ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117124ZHE#01

Quant Time: Feb 16 00:17:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	659143	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1120445	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1057274	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	418154	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	353535	53.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.92%	
35) Dibromofluoromethane	7.59	113	354365	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
50) Toluene-d8	10.09	98	1362823	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
62) 4-Bromofluorobenzene	12.40	95	452535	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
Target Compounds						
16) Acetone	3.82	43	14494	8.545	ug/l	98
20) Methylene Chloride	4.56	84	10611	1.491	ug/l	93
43) Isopropyl Acetate	8.17	43	32644	2.997	ug/l #	82
80) 1,3,5-Trimethylbenzene	12.73	105	7515	0.308	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021519\
Data File : VN053810.D
Acq On : 15 Feb 2019 19:46
Operator : JC/SP
Sample : PB117124ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117124ZHE#01

Quant Time: Feb 16 00:17:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

