

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021519\
 Data File : VN053819.D
 Acq On : 16 Feb 2019 00:40
 Operator : JC/SP
 Sample : PB117124ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117124ZHE#04

Manual Integrations
 APPROVED

MMDadoda
 2/18/2019 9:48:29 AM

Quant Time: Feb 16 05:54:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	661921	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1153328	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1005202	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	382357	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	364746	54.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.86%	
35) Dibromofluoromethane	7.59	113	369296	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
50) Toluene-d8	10.09	98	1307811	47.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.58%	
62) 4-Bromofluorobenzene	12.40	95	425705	46.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.30%	
Target Compounds						
16) Acetone	3.82	43	17318	10.168	ug/l	91
20) Methylene Chloride	4.56	84	11771m	1.647	ug/l	
43) Isopropyl Acetate	8.17	43	32896	2.934	ug/l #	84
80) 1,3,5-Trimethylbenzene	12.73	105	9829	0.441	ug/l	89
95) Naphthalene	15.13	128	5139	2.032	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021519\
Data File : VN053819.D
Acq On : 16 Feb 2019 00:40
Operator : JC/SP
Sample : PB117124ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117124ZHE#04

Manual Integrations
APPROVED

MMDadoda
2/18/2019 9:48:29 AM

Quant Time: Feb 16 05:54:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

