

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021519\
 Data File : VN053820.D
 Acq On : 16 Feb 2019 1:07
 Operator : JC/SP
 Sample : PB117124ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117124ZHE#05

Quant Time: Feb 16 05:57:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	691525	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1159362	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1159562	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	429497	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	366983	52.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.80%	
35) Dibromofluoromethane	7.59	113	365293	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	10.09	98	1405041	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	12.40	95	464306	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
Target Compounds						
16) Acetone	3.82	43	16238	9.125	ug/l	98
20) Methylene Chloride	4.56	84	10413	1.395	ug/l	88
43) Isopropyl Acetate	8.17	43	33005	2.928	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021519\
Data File : VN053820.D
Acq On : 16 Feb 2019 1:07
Operator : JC/SP
Sample : PB117124ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117124ZHE#05

Quant Time: Feb 16 05:57:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

