

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021519\
 Data File : VN053825.D
 Acq On : 16 Feb 2019 3:21
 Operator : JC/SP
 Sample : PB117124ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117124ZHE#10

Quant Time: Feb 18 08:19:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	641953	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1079382	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1018076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	399668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	344456	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
35) Dibromofluoromethane	7.59	113	343626	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
50) Toluene-d8	10.09	98	1321187	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
62) 4-Bromofluorobenzene	12.40	95	436094	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
Target Compounds						
16) Acetone	3.82	43	19258	11.658	ug/l	90
20) Methylene Chloride	4.55	84	21622	3.120	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021519\
Data File : VN053825.D
Acq On : 16 Feb 2019 3:21
Operator : JC/SP
Sample : PB117124ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117124ZHE#10

Quant Time: Feb 18 08:19:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

