

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021519\
 Data File : VN053827.D
 Acq On : 16 Feb 2019 4:14
 Operator : JC/SP
 Sample : PB117124ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117124ZHE#12

Quant Time: Feb 18 08:22:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	642637	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1086932	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1028031	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	426281	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	357928	55.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.04%	
35) Dibromofluoromethane	7.59	113	348114	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
50) Toluene-d8	10.09	98	1329746	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
62) 4-Bromofluorobenzene	12.40	95	445508	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
Target Compounds						
16) Acetone	3.82	43	14720	8.902	ug/l	91
20) Methylene Chloride	4.56	84	7826	1.128	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021519\
Data File : VN053827.D
Acq On : 16 Feb 2019 4:14
Operator : JC/SP
Sample : PB117124ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117124ZHE#12

Quant Time: Feb 18 08:22:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

