


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021519\
 Data File : VN053789.D
 Acq On : 15 Feb 2019 9:12
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Title : SW846 8260
 Last Update : Fri Feb 01 09:11:52 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	7552	PASS
75	95	30	60	46.7	21645	PASS
95	95	100	100	100.0	46333	PASS
96	95	5	9	6.7	3094	PASS
173	174	0.00	2	1.4	606	PASS
174	95	50	100	93.6	43352	PASS
175	174	5	9	7.8	3380	PASS
176	174	95	101	98.2	42592	PASS
177	176	5	9	6.9	2927	PASS