

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021523\
 Data File : VN076694.D
 Acq On : 15 Feb 2023 18:52
 Operator : JC\MD
 Sample : 01547-03 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-3

Quant Time: Feb 16 04:35:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N021523W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Feb 16 01:08:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.830	128	26716	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	9.107	114	176500	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	190037	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	81217	33.238	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	110.800%#
60) 4-Bromofluorobenzene	12.854	95	83157	24.739	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.467%
63) Toluene-d8	10.571	98	201514	30.274	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.900%
Target Compounds						
25) Chloroform	7.978	83	58045	10.140	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021523\
Data File : VN076694.D
Acq On : 15 Feb 2023 18:52
Operator : JC\MD
Sample : 01547-03 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-3

Quant Time: Feb 16 04:35:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N021523W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Feb 16 01:08:46 2023
Response via : Initial Calibration

