

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021524\
 Data File : VN081037.D
 Acq On : 15 Feb 2024 14:01
 Operator : JC\MD
 Sample : VN0215MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0215MBL01

Quant Time: Feb 16 00:24:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 07:59:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	391295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	816985	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	962705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	403228	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	236028	49.040	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.080%
35) Dibromofluoromethane	8.165	113	210081	44.849	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.700%
50) Toluene-d8	10.571	98	836004	50.606	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.220%
62) 4-Bromofluorobenzene	12.847	95	393944	70.116	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	140.240%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021524\
Data File : VN081037.D
Acq On : 15 Feb 2024 14:01
Operator : JC\MD
Sample : VN0215MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0215MBL01

Quant Time: Feb 16 00:24:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 07:59:19 2024
Response via : Initial Calibration

