

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021524\
 Data File : VN081038.D
 Acq On : 15 Feb 2024 14:24
 Operator : JC\MD
 Sample : VN0215MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0215MBL02

Quant Time: Feb 16 00:24:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 07:59:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	569423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	1022830	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	922312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	323474	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	336509	48.045	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.100%
35) Dibromofluoromethane	8.165	113	281017	47.919	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.840%
50) Toluene-d8	10.565	98	985208	47.635	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.280%
62) 4-Bromofluorobenzene	12.847	95	341160	48.501	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021524\
Data File : VN081038.D
Acq On : 15 Feb 2024 14:24
Operator : JC\MD
Sample : VN0215MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0215MBL02

Quant Time: Feb 16 00:24:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 07:59:19 2024
Response via : Initial Calibration

